

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

FEB 16 2010

Supervisor/Operator Performing the Verification Procedure:		DataMaster cdm S/N	130143 ✓
Name	WADE L TULLIVER	ID#	3805
Date	11 FEB 10		
A	Agency	FT WAINWRIGHT POLICE	Phone # 353-7553
	Instrument Location	Bldg 30 28 ✓	
Alco S/N		X 301556 ✓	Target Value .083 ✓
			High Pressure 500 ²
B	Alco Test Values	.080 : .087 ✓	.085 ✓
	1 st Alco	2 nd Alco	
Signature		✓	
(OVER)			

(Do Not write in the area below)


I, Nita J. Bolz, after being first duly sworn, depose and state as follows:

- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
- (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.
- (6) The referenced instrument is certified for evidentiary use in the State of Alaska.

Nita J. Bolz
Scientific Director
State Breath Alcohol Program

Subscribed and sworn before me this 23rd day of February, 2010.

Carolyn M. Noland
Notary Public, State of Alaska
Commission Expires with Office



VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130143 ✓

(CONTINUED FROM FRONT PAGE)

C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

~~NONDRINKING~~ SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130143 ✓

FEBRUARY 11, 2010

OPERATOR'S NAME:

TOLLIVER, MADE L

OPERATOR'S NUMBER: 3004

SUBJECT'S LAST NAME:

DOE

SUBJECT'S FIRST NAME/MI:

JOHN

O.L. #: 00000000

DEPT/AGENCY: FAI4

CASE/REPORT: 00-0000

TEST TYPE: T

ALCO TARGET VALUE: .083 ✓

ALCO S/N: X301556 ✓

— BREATH ANALYSIS —

.083 ADJUSTED FOR 29.08 in	✓
ALCO TARGET	.080 14:10
BLANK TEST	.000 14:11
INTERNAL STANDARD	VERIFIED 14:11
ALCO TV 29.08 in	.087 14:12
BLANK TEST	.000 14:12
SUBJECT SAMPLE	.000 14:13
BLANK TEST	.000 14:14
ALCO TV 29.08 in	.085 14:14
BLANK TEST	.000 14:15

DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130143 ✓

FEBRUARY 11, 2010 ✓
TIME 14:58

— DIAGNOSTIC CHECK —

COMPUTER:	OKAY
PROGRAM:	OKAY
SOFTWARE DATE:	02/20/01
HEATERS	
SAMPLE CHAMBER:	49c
BREATH TUBE:	41c
BAROMETER:	29.07 in
FLOW DETECTOR:	OKAY
PUMP	
HIGH SPEED:	OKAY
DETECTOR:	OKAY
FILTERS:	OKAY
QUARTZ STANDARD:	OKAY
CALIBRATION:	OKAY

PRINTER TEST

"#12"()+-./0123456789:;<=>@#B00EF6
HIJKL NOPQRSTU VWXYZ\]^_`ab cde fghijklmno
pqrstu vwxyz{|}~